

QUALITY ASSURANCE / QUALITY CONTROL DATA

J

Preparation Batch ID: 012117BTKN				Analysis Method: EPA 351.2				Preparation Type: 351.2				
Method Batch ID: M012117BTKN								Preparation Date:				
Analyte	MDL	PQL	Result	Qual	Units	Spike Amount	% REC	% REC Low Limit	-	% REC High Limit	%RPD	% RPD Limit
QA/QC Type: MB Lab Sample ID: 012117BTKNMB				Client Sample ID: 012117BTKNMB				Date Analyzed: 1/21/2017 10:53:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	0.24	U	mg/L							
QA/QC Type: LCS Lab Sample ID: 012117BTKNLCS				Client Sample ID: 012117BTKNLCS				Date Analyzed: 1/21/2017 10:46:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	4.26		mg/L	4.50	94.7	90	-	110		
QA/QC Type: LCSD Lab Sample ID: 012117BTKNLCSD				Client Sample ID: 012117BTKNLCSD				Date Analyzed: 1/21/2017 10:48:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	4.29		mg/L	4.50	95.3	90	-	110	0.70	20
QA/QC Type: MS Lab Sample ID: 012117BTKNMS				Client Sample ID: 207934MS				Date Analyzed: 1/21/2017 11:01:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	5.86		mg/L	5.00	104	90	-	110		
QA/QC Type: MSD Lab Sample ID: 012117BTKNMSD				Client Sample ID: 207934MSD				Date Analyzed: 1/21/2017 11:02:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	5.86		mg/L	5.00	104	90	-	110	0	20
QA/QC Type: DUP Lab Sample ID: 012117BTKNDUP				Client Sample ID: 207934DUP				Date Analyzed: 1/21/2017 11:00:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	0.64		mg/L						0	20

Comments:

Preparation Batch ID: 012317BNO23				Analysis Method: EPA 353.2 (Nitrate (N))				Preparation Type: Gen Prep				
Method Batch ID: M012317BNO23								Preparation Date: 1/23/2017 11:03:00 AM				
Analyte	MDL	PQL	Result	Qual	Units	Spike Amount	% REC	% REC Low Limit	-	% REC High Limit	%RPD	% RPD Limit
QA/QC Type: MB Lab Sample ID: 012317BNO23MB				Client Sample ID: 012317BNO23MB				Date Analyzed: 1/23/2017 11:03:00 AM				
Nitrate (N)	0.042	0.050	0.042	U	mg/L							
Nitrate-Nitrite (N)	0.042	0.050	0.042	U	mg/L							
QA/QC Type: LCS Lab Sample ID: 012317BNO23LCS				Client Sample ID: 012317BNO23LCS				Date Analyzed: 1/23/2017 10:57:00 AM				
Nitrate-Nitrite (N)	0.042	0.050	2.04		mg/L	2.04	100	90	-	110		
Nitrate (N)	0.042	0.050	2.04		mg/L	2.04	100	90	-	110		
QA/QC Type: LCSD Lab Sample ID: 012317BNO23LCSD				Client Sample ID: 012317BNO23LCSD				Date Analyzed: 1/23/2017 10:58:00 AM				