

QUALITY ASSURANCE / QUALITY CONTROL DATA



Preparation Batch ID: 020817BTP
Method Batch ID: M020817BTP

Analysis Method: EPA 365.1 (Phosphorus -Total)

Preparation Type: 365.1
Preparation Date:

Analyte	MDL	PQL	Result	Qual	Units	Spike Amount	% REC	% REC Low Limit	-	% REC High Limit	%RPD	% RPD Limit
QA/QC Type: MB Lab Sample ID: 020817BTPMB				Client Sample ID: 020817BTPMB				Date Analyzed: 2/8/2017 4:14:00 AM				
Phosphorus- Total	0.023	0.050	0.023	U	mg/L							
QA/QC Type: LCS Lab Sample ID: 020817BTPLCS				Client Sample ID: 020817BTPLCS				Date Analyzed: 2/8/2017 4:07:00 AM				
Phosphorus- Total	0.023	0.050	2.37		mg/L	2.35	101	90	-	110		
QA/QC Type: LCSD Lab Sample ID: 020817BTPLCSD				Client Sample ID: 020817BTPLCSD				Date Analyzed: 2/8/2017 4:08:00 AM				
Phosphorus- Total	0.023	0.050	2.38		mg/L	2.35	101	90	-	110	0.42	20
QA/QC Type: MS Lab Sample ID: 020817BTPMS				Client Sample ID: 208723MS				Date Analyzed: 2/8/2017 4:24:00 AM				
Phosphorus- Total	0.023	0.050	1.19		mg/L	1.00	106	90	-	110		
QA/QC Type: MSD Lab Sample ID: 020817BTPMSD				Client Sample ID: 208723MSD				Date Analyzed: 2/8/2017 4:26:00 AM				
Phosphorus- Total	0.023	0.050	1.19		mg/L	1.00	106	90	-	110	0	20
QA/QC Type: DUP Lab Sample ID: 020817BTPDUP				Client Sample ID: 208723DUP				Date Analyzed: 2/8/2017 4:22:00 AM				
Phosphorus- Total	0.023	0.050	0.13		mg/L						0	20

Comments:

Preparation Batch ID: 021017ATKN
Method Batch ID: M021017ATKN

Analysis Method: EPA 351.2

Preparation Type: 351.2
Preparation Date:

Analyte	MDL	PQL	Result	Qual	Units	Spike Amount	% REC	% REC Low Limit	-	% REC High Limit	%RPD	% RPD Limit
QA/QC Type: MB Lab Sample ID: 021017ATKNMB				Client Sample ID: 021017ATKNMB				Date Analyzed: 2/10/2017 10:26:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	0.24	U	mg/L							
QA/QC Type: LCS Lab Sample ID: 021017ATKNLCS				Client Sample ID: 021017ATKNLCS				Date Analyzed: 2/10/2017 10:19:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	6.63		mg/L	7.15	92.7	90	-	110		
QA/QC Type: LCSD Lab Sample ID: 021017ATKNLCSD				Client Sample ID: 021017ATKNLCSD				Date Analyzed: 2/10/2017 10:21:00 AM				
Nitrogen- Total Kjeldahl	0.24	0.50	6.65		mg/L	7.15	93.0	90	-	110	0.30	20