

QUALITY ASSURANCE / QUALITY CONTROL DATA



Preparation Batch ID: 020617ANO23			Analysis Method: EPA 353.2 (Nitrate-Nitrite (N))				Preparation Type: Gen Prep				
Method Batch ID: M020617ANO23			Preparation Date: 2/6/2017 10:55:00 AM								
Analyte	MDL	PQL	Result	Qual	Units	Spike Amount	% REC	% REC Low Limit	% REC High Limit	%RPD	% RPD Limit

QA/QC Type: DUP	Lab Sample ID: 020617ANO23DUP	Client Sample ID: 208244DUP	Date Analyzed: 2/6/2017 11:00:00 AM								
Nitrate-Nitrite (N)	0.042	0.050	0.86	mg/L						1.2	20
Nitrate (N)	0.042	0.050	0.86	mg/L							20

Comments:

Preparation Batch ID: 020717BTKN			Analysis Method: EPA 351.2				Preparation Type: 351.2				
Method Batch ID: M020717BTKN			Preparation Date:								
Analyte	MDL	PQL	Result	Qual	Units	Spike Amount	% REC	% REC Low Limit	% REC High Limit	%RPD	% RPD Limit

QA/QC Type: MB	Lab Sample ID: 020717BTKNMB	Client Sample ID: 020717BTKNMB	Date Analyzed: 2/7/2017 10:58:00 PM								
Nitrogen- Total Kjeldahl	0.24	0.50	0.24	U	mg/L						

QA/QC Type: LCS	Lab Sample ID: 020717BTKNLCS	Client Sample ID: 020717BTKNLCS	Date Analyzed: 2/7/2017 10:51:00 PM								
Nitrogen- Total Kjeldahl	0.24	0.50	6.63		mg/L	7.15	92.7	90	-	110	

QA/QC Type: LCSD	Lab Sample ID: 020717BTKNLCS	Client Sample ID: 020717BTKNLCS	Date Analyzed: 2/7/2017 10:52:00 PM								
Nitrogen- Total Kjeldahl	0.24	0.50	6.68		mg/L	7.15	93.4	90	-	110	0.75

QA/QC Type: MS	Lab Sample ID: 020717BTKNMS	Client Sample ID: 208314MS	Date Analyzed: 2/7/2017 11:06:00 PM								
Nitrogen- Total Kjeldahl	0.24	0.50	5.03		mg/L	5.00	92.2	90	-	110	

QA/QC Type: MSD	Lab Sample ID: 020717BTKNMSD	Client Sample ID: 208314MSD	Date Analyzed: 2/7/2017 11:07:00 PM								
Nitrogen- Total Kjeldahl	0.24	0.50	5.03		mg/L	5.00	92.2	90	-	110	0

QA/QC Type: DUP	Lab Sample ID: 020717BTKNDUP	Client Sample ID: 208314DUP	Date Analyzed: 2/7/2017 11:04:00 PM								
Nitrogen- Total Kjeldahl	0.24	0.50	0.42	I	mg/L					0	20

Comments: